



The company Dux B.V., Zonnebaan 14, NL-3542 EC Utrecht, declares that the **PeelVue+™ Sterilizing Pouches**, purpose of this declaration, comply with the guidelines mentioned below:

Council Directive 93/42/EEC – of 14 June 1993 (and annexes) concerning medical devices
ISO 11607 – packaging for terminally sterilized medical devices

Dux B.V. states that, with regard to maintaining product sterility, there is no difference in performance between the 10mm self adherent seal of PeelVue+™ Sterilizing Pouches and the 8mm heat seal obtained by a sealing machine.

Dux B.V.

Mr. J. Doorman
Managing Director



Zonnebaan 14
3542 EC Utrecht
The Netherlands